



Ref. Certif. No.

DK-14043-A2

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product
Produit

Switch Mode Power Supply Module

Name and address of the applicant
Nom et adresse du demandeur

**VOX POWER LTD
TERENURE ENTERPRISE CENTRE
17 RATHFARNHAM ROAD TERENURE
DUBLIN 6W IRELAND
VOX POWER LTD
TERENURE ENTERPRISE CENTRE
17 RATHFARNHAM ROAD TERENURE
DUBLIN 6W IRELAND
CELETRONIX POWER (INDIA) PVT LTD
UNIT 41 SDF-II
SEEPZ ANDHERI (E)
MUMBAI MH 400096 INDIA**

Name and address of the manufacturer
Nom et adresse du fabricant

Name and address of the factory
Nom et adresse de l'usine

*Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{me} page*

Rating: See appendix, IPX0

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Trademark (if any)
Marque de fabrique (si elle existe)

Vox-Power



Model / Type Ref.
Ref. De type

NEVO600M-abcd-zzzz or NEVO+600M-abcd-zzzz, See appendix

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire, peuvent être indiqués sur la 2^{me} page)

See appendix

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

PUBLICATION	EDITION
IEC 60601-1:1988 + A1:1991 + A2:1995	2nd

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

**E316486-A4-CB-1 with Amendment 2 issue date 2009-11-04
with Amendment 3 issue date 2010-01-07**

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



UL International Demko A/S
Lyskaer 8, P.O. Box 514, DK-2730 Herlev, Denmark
Tel: +45 44856565, Fax: +45 44856500

Date:
2010-02-05

Signature:
Jan-Erik Storgaard
Certification manager



Ref. Certif. No.

DK-14043-A2

Model:

NEVO600M-abcd-zzzz or NEVO+600M-abcd-zzzz

where:

a (slot A)=0,1,2, 3, 4 or 5;

b (slot B)=0,1,2, 3, 4 or 5;

c (slot C)=0,1,2, 3, 4 or 5;

d (slot D)=0,1,2, 3, 4 or 5; and where:

0=No Module (blanking plate fitted); 1=OP1, 2=OP2; 3=OP3, 4=OP4, 5=OP5.

May also be followed by up to 4 alphanumeric zzzz indicating non-safety related information.

Output Modules:

OP1; OP2; OP3; OP4; OP5.

Rating(s):

NEVO600M-abcd-zzzz: 100-240Vac, 4.5A, 50-60Hz and 120-300Vdc, 4.5A

NEVO+600M-abcd-zzzz: 100-240Vac, 6A, 50-60Hz and 120-300Vdc, 6A

Maximum output power 600W. Individual outputs rated according to modules fitted.

Output Module Output Ratings:

OP1=1.5-7.5Vdc, 25A, 125W.

OP2=4.5-15Vdc, 15A, 150W.

OP3=9-30Vdc, 7.5A, 150W.

OP4=18-58Vdc, 3.75A, 150W.

OP5=5-15Vdc, 5A, 150W and/or 5-15Vdc, 5A, 150W.

(OP5 may have either 1 or 2 outputs)

Additional Information:

This is an amendment to DK-14043-A1 issued 2009-02-04 due to amendment of Model Number nomenclature, deleting some models, removing Schematics from Enclosures; update of trademark, change of manufacturing locations, amendment of critical components table and alternate construction.

Also investigated to EN 60601-1: 1990 + A1:1993 + A2:1995

Additional information (if necessary)

Information complémentaire (si nécessaire)



Date:

2010-02-05

UL International Demko A/S

Lyskaer 8, P.O. Box 514, DK-2730 Herlev, Denmark
Tel: +45 44856565, Fax: +45 44856500

Signature:

Jan-Erik Storgaard
Certification manager