

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product	AC/DC switch mode power supply
Name and address of the applicant	VOX POWER LTD Unit 2, Redcow Interchange Estate Ballymount Dublin 22, D22 Y8H2 IRELAND
Name and address of the manufacturer	VOX POWER LTD Unit 2, Redcow Interchange Estate Ballymount Dublin 22, D22 Y8H2 IRELAND
Name and address of the factory	Panyu Trio Microtronics Co Ltd SHIJI INDUSTRIAL ESTATE DONGYONG NANSHA GUANGZHOU GUANGDONG, 511453 China
Note: When more than one factory, please report on page 2	<input checked="" type="checkbox"/> Additional Information on page 2
Ratings and principal characteristics	Input: 100 - 240 Vac, 50-60 Hz, 3A Output: 12Vdc to 58Vdc (depends on model), max. 300 W For mains and thermal deratings please see Additional Information section in Test Report
Trademark / Brand (if any)	VOX-POWER
Type of Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	VCCS300M-aaaa-bbb See Page 2
Additional information (if necessary may also be reported on page 2)	National Differences specified in the CB Test Report. <input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005
As shown in the Test Report Ref. No. which forms part of this Certificate	E316486-D1001-1/A0/C0-CB issued on 2020-12-16

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pflugsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2020-12-18

Signature:

Jan-Erik Storgaard

Model Details:

VCCS300M-aaaa-bbb - aaaa denotes the output voltage and may be any value from 12 to 58 (including decimals to one significant digit);
- bbb may be blank or alphanumeric characters;

Factories:

Trio-Tronics (Thailand) Ltd.
7/295 Mu. 6, Map Yang Phon Sub-District
Pluak Daeng District, Rayong Province Thailand

Additional Information:

The risk management requirements of the standard were not addressed.
Additionally evaluated to EN 60601-1:2006 / A1:2013 / A12:2014.

Additional information (if necessary)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2020-12-18

Signature:


Jan-Erik Storgaard