



Ref. Certif. No.

DK-113449-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product Switch mode modular power supplies for installation into medical equipment.

Name and address of the applicant VOX POWER LTD
UNIT 2
REDCOW INTERCHANGE ESTATE DUBLIN 22
IRELAND

Name and address of the manufacturer VOX POWER LTD
UNIT 2
REDCOW INTERCHANGE ESTATE DUBLIN 22
IRELAND

Name and address of the factory PANYU TRIO MICROTRONICS CO LTD
SHIJI INDUSTRIAL ESTATE
DONGYONG
NANSHA GUANGZHOU, GUANGDONG, 511453
CHINA

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

[Additional Information on page 2](#)
 Input: 100-240 V ac or 120-370 V dc, 50-60Hz , 7A
 Output: 1.5-7.5V, 4.5-15V, 9-30V, 18-58V, 25A, 15A, 7.5A, 3.75A, 125W, 150W, 150W, 150W, DC , DC , DC , DC ; Insulation Class: Class I

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

Model / Type Ref. VCCM600M (i)
 [Additional Information on page 2](#)

Additional information (if necessary may also be reported on page 2)

The risk management requirements of the standard were not addressed
 [Additional Information on page 2](#)

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012

As shown in the Test Report Ref. No. which forms part of this Certificate

E316486-D1000-2/A0/C0-ULCBINF issued on 2021-05-14

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2021-05-14

Signature: Jan-Erik Storgaard



Ref. Certif. No.

DK-113449-UL

Factories:

TRIO-TRONICS (THAILAND) LTD
7/295 MU. 6
MAP YANG PHON SUB-DISTRICT PLUAK DAENG DISTRICT, RAYONG PROVINCE,
THAILAND

Additional Model(s):

Series: VCCM600M (i), (i) may be blank, or comprised of VCCM600M which may be followed by: 1, 2, 3 or 4, and maybe followed by alphanumeric characters to indicate non-safety related information.

Where:

1 (Slot 1) = 0, A, B, C or D

2 (Slot 2) = 0, A, B, C or D

3 (Slot 3) = 0, A, B, C or D

4 (Slot 4) = 0, A, B, C or D

Where 0 = No output module fitted, A = Module OPA, B = Module OPB, C = Module OPC, D = Module OPD

Additionally evaluated to:

EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014
National Differences specified in the CB Test Report.

Additional information (if necessary)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2021-05-14

Signature:

Jan-Erik Storgaard